

Document Title: <b>Description, Leaf Porometer Sensor Head Replacement Instructions</b>		Part # and Rev. <b>13956</b>	
		Release Date: <b>1.8.2010</b>	
<b>Rev.</b>	<b>Description</b>	<b>Revision By</b>	<b>Date</b>
-02	Update new sensor, image	DDH	12.15.2011
03	Edited according to DCO 04245	CSC	7.17.2019

Production File path: [http://publications.decagon.com/Inserts/13956\\_SC-1%20Sensor%20Head%20Replacement%20Instructions\\_Print.pdf](http://publications.decagon.com/Inserts/13956_SC-1%20Sensor%20Head%20Replacement%20Instructions_Print.pdf)

Working File Path: [https://drive.google.com/drive/u/1/folders/1L6iz\\_x5fhlauiPzSvougQLpff0qEvUt](https://drive.google.com/drive/u/1/folders/1L6iz_x5fhlauiPzSvougQLpff0qEvUt)

**Dimensions:** print size 8.5" wide x 11" tall, cut in half, final size is 8.5" wide x 5.5" tall

**Colors:** CMYK/Full color 4/4

**Printer:** In-house printing

**Material:** Card stock, Bright White, 216 g/m<sup>2</sup> (folding cardstock)

**Finishing:** NA

Illustrations are Ref Only \*\* Not to Scale \*\*




## SC-1 POROMETER SENSOR HEAD REPLACEMENT INSTRUCTIONS

13956-03  
7.17.2019

### METER

Before using the new SC-1 sensor head, follow the directions below to save the new sensor head serial number and default calibration number into the SC-1 controller. Ensure the SC-1 has the latest firmware update before beginning this procedure ([metergroup.com/sc1-support](http://metergroup.com/sc1-support)).

1. Connect the new sensor head to the controller.
2. Locate the sensor head serial number and calibration number on the tag attached near the serial connector on the sensor head cable.
3. Turn on the SC-1 controller.
4. Hold down **Enter** and **ESC** simultaneously for several seconds until the Factory Calibration screen appears.
5. Press **DOWN** to highlight Sensor and press **Enter**.
6. Use the **UP** and **DOWN** to enter the new serial number. Press **Enter** to advance through each value in turn and to return to the Factory Calibration screen.
7. Press **DOWN** to highlight Cal Num and press **Enter**.
8. Use the **UP** and **DOWN** to enter the new serial number. Press **Enter** to advance through each value in turn and to return to the Factory Calibration screen.
9. Press **DOWN** to highlight Quit Factory Mode and press **Enter**.



The correct serial number and default calibration number are now in the SC-1 controller memory. The default calibration number will be applied until a calibration is performed or when the default calibration is restored. METER recommends that the accuracy of the measurement be verified each day the SC-1 is used or every time it is used under a different set of environmental conditions (refer to the SC-1 User Manual for details on accuracy verification). A user-initiated calibration should be performed if the accuracy is outside of the 10% specification.

If the old sensor head was not equipped with the black desiccant chamber, download the SC-1 User Manual ([metergroup.com/sc1-support](http://metergroup.com/sc1-support)) and review the operation of the desiccant-equipped sensor head.

**METER Group, Inc. USA**  
2365 NE Hopkins Court, Pullman, WA 99163  
T +1.509.332.2756 F +1.509.332.5158  
E [info@metergroup.com](mailto:info@metergroup.com) W [metergroup.com](http://metergroup.com)

**NORTH AMERICA**  
E [support.environment@metergroup.com](mailto:support.environment@metergroup.com)  
T +1.509.332.5600

**EUROPE**  
E [support.europe@metergroup.com](mailto:support.europe@metergroup.com)  
T +49 89 12 66 52 0